

BIO-CLEAN NE

Neutral Detergent

Bio-Clean NE is highly recommended for all scopes and cannulated instruments where Biofilm build-up can develop as a result of constant exposure to wet and dry phases during usage and reprocessing.

Perfect for manual reprocessing

Cleans, washes and prepares semi-critical medical devices enabling high level disinfectants to kill and remove biofilm.

The only detergent designed to target and dissolve polysaccharides from semi-critical and critical medical devices.

Getting rid of biofilm is a two-step process; mechanical cleaning action using Bio-Clean NE, followed by the use of a liquid chemical sterilant or high level disinfectant. Is the simplest, most inexpensive method of breaking down biofilm and killing the underlying bacteria.

Features and Benefits:

- Removes all bio-burden – blood, carbohydrates, protein, polysaccharides, fats, oils, uric acid and other nitrogenous compounds.
- Cleans the inanimate surfaces where biofilm, germs, allergens or micro-organisms can hide, thrive and grow.
- Biologically active additives speed the process of liquefaction and solubilisation.
- Compatible with all medical device material.
- Low sudsing, Neutral pH, Non-Abrasive, Free Rinsing and 100% Biodegradable.
- Works in all water temperatures.

Available in:

Cat No. 345BNE 4 x 4 Litres

NHSSC Code:

FAL14142



Directions for use:

1. Dilute 4ml per litre of water.
2. Submerge instruments and scopes to be cleaned. For scopes: Suction or flush through channels before soaking.
3. Soak for 2 minutes to remove all organic soils (Heavy soil may require longer time).
4. Rinse thoroughly with tap, distilled or sterile water.
5. Discard used solution.

Cautions

Avoid prolonged contact with skin. Do not swallow. Keep away from children